

## Certificate of Analysis

**Product Description:** TBE 10X Buffer, Molecular Biology Grade\_  
**Catalog Number:** TBE10X02\_  
**Lot:** 143600\_

**Formula:** N/A  
**CAS #:** N/A  
**Physical Description:** Clear Colorless Solution

**Formula Weight:** N/A  
**Storage:** 15 - 30°C

Test	Specification	Results
Identity Test	Passes	Passes

pH 1X conc.(25°C): 8.41  
pH 10X conc.(25°C): 8.24  
Conductivity 1X conc. (25°C): 1.16 mS/cm  
Conductivity of 10X conc.(25°C): 4.96 mS/cm  
RNase: Not detected  
DNase: Not detected  
Protease: Not detected  
Recommended Retest Date (MM/DD/YYYY): 07/20/2022



07/21/2020 - John Huang, PhD  
MP Biomedicals, LLC.  
Quality Control Manager

This is an electronically generated document  
<mailto:biotech@mpbio.com>  
<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site  
Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)