

## Certificate of Analysis

**Product Description:** SSC 20x Buffer\_  
**Catalog Number:** RIST2050\_  
**Lot:** SR02136\_

**Formula:** N/A  
**CAS #:** N/A  
**Physical Description:** Solution

**Formula Weight:** N/A  
**Storage:** 15 - 30°C

Test	Specification	Result
Identity Test	Passes	Passes
Expiration Date	Month/Day/Year	06/29/2021

pH of 1X conc.: 7.8  
Conductivity of 20X conc.: 162 mS/cm  
RNase: None detected



12/30/2019 - John Huang, PhD  
MP Biomedicals, LLC.  
Quality Control Manager

This is an electronically generated document  
<mailto:biotech@mpbio.com>  
<http://www.mpbio.com>

Online Ordering, MSDSs, certificates of analysis and data sheets now available on our web site  
Technical Service: 1-800-279-5490 (440-337-1200) Customer Service: 1-800-854-0530 (440-337-1200)